

INSTRUCTIONS FOR USE

DESCRIPTION OF THE PRODUCT

The One-Step Centesis Catheter is a short-term drainage catheter, available in straight, pigtail, valved and non-valved configurations with introducer needle.

INDICATIONS FOR USE

The One-Step Centesis Catheter is indicated for the percutaneous drainage of fluids.

CONTRAINDICATIONS

None known.

PRECAUTIONS

- Read manufacturer's instructions prior to use
- **Rx Only Caution:** Federal (USA) law restricts this device to sale by or on the order of a physician trained and/or experienced in the use of the device.
- Contents of unopened, undamaged package are sterile.
- Do not leave in patient longer than 24 hours.
- Store in a cool, dry place.
- Do not use if packaging is damaged.
- Inspect the device prior to use to verify that no damage has occurred during shipping.
- Do not use after expiration date
- For single patient use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or reesterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or reesterilization may also create a risk of contamination of the device and/or cause patient infection or crossinfection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

ADVERSE REACTIONS

- Pneumothorax
- Bleeding complications
- Re-expansion pulmonary edema

INSTRUCTIONS FOR USE- PIGTAIL CATHETER

1. Remove the needle from the protective boot.
2. Insert the needle into catheter.
3. Insert the needle through the catheter until the needle hub is fully seated in the catheter hub.
4. Remove the distal obturator from the needle prior to inserting catheter into patient. (see Figure 1)




Figure 1

Warning: Do not re-attach distal obturator to needle.

5. Insert catheter into fluid collection site using standard insertion technique.
 6. When catheter is in appropriate position, remove needle from catheter.
- Warning:** Do not reinsert needle into catheter while the catheter is still in the patient.
7. Attach a device with a standard Luer fitting to catheter hub to remove fluid.
 8. Following the procedure, gently withdraw the catheter and dress the site accordingly.

INSTRUCTIONS FOR USE- STRAIGHT CATHETER

1. Remove the catheter and needle from protective boot(s).
- Note:** Steps 2 and 3 below are only for 6F and 8F Valved One-Step Centesis Catheters.
2. Insert the needle through the catheter until the needle hub is fully seated in the catheter hub.
 3. Remove the distal obturator from the needle prior to inserting catheter into patient. (see Figure 2)




Figure 2

Warning: Do not re-attach distal obturator to needle.

4. Insert catheter into fluid collection site using standard insertion technique.
 5. When catheter is in appropriate position, remove needle from catheter.
- Warning:** Do not reinsert needle into catheter while the catheter is still in the patient.
6. Attach a device with a standard Luer fitting to catheter hub to remove fluid.
 7. Following the procedure, gently withdraw the catheter and dress the site accordingly.



MERITMEDICAL®

CE 0086

**Manufacturer:**

Merit Medical Systems, Inc. 1600 W. Merit Pkwy
South Jordan, Utah 84095
U.S.A. 1-801-253-1600
U.S.A. Customer Service 1-800-356-3748

www.merit.com

EC REP

Authorized Representative:

Merit Medical Ireland Ltd, Parkmore Business Park West,
Galway, Ireland

EC Customer Service +31 43 3588222