


AEROmini[®]

IMPLANTED DEVICE PATIENT INFORMATION CARD



 **MERTMEDICAL**
ENDOTEK[®]

AERO*mini*[®]

Fully Covered Tracheobronchial Stent

Patient Name: _____ Stent Lot #: _____

Date of Implant: _____ Stent Size: _____

Implant Location: _____

Implanting Physician: _____

Hospital: _____ Contact Number: _____



MR Conditional

MRI Safety Information

Non-clinical testing has demonstrated that the AEROMini is MR Conditional. A patient with this device can be safely scanned in an MR system meeting the following conditions:

- Static magnetic field of 1.5T and 3.0T only
- Maximum spatial gradient magnetic field of 3,000 Gauss/cm (30 T/m) or less
- Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2 W/kg (Normal Operating Mode)

Under the scan condition defined above, the AEROMini is expected to produce a maximum temperature rise of 2.0°C after 15 minutes of continuous scanning.

In non-clinical testing, the image artifact caused by the device extends approximately 5 mm from the AEROMini when imaged with a gradient echo pulse sequence and a 3Tesla MRI system. The artifact does obscure the device lumen.



EXPANDING THE POSSIBILITIES™

www.merit.com/endotek

Merit Medical Systems, Inc. • 1600 West Merit Parkway • South Jordan, Utah 84095 • 1-801-253-1600 • 1-800-35-MERIT

403058001/C ID 102317