

CLINICAL BENEFITS

- Dual and single port plugs to fit most bipolar generators
- Hydrophilic coated tip to reduce adherence
- Silver electrode for optimal conductivity

QUICK FACTS

- Box quantity: 1
- Single use
- Not made with natural rubber latex or DEHP
- Sterility: ETO
- Shelf life: 3 years
- FDA Class II medical device
 - 510 (K) number: K102273
 - <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMN.pmn.cfm?ID=K102273>

CODING*

Healthcare Common Procedure Code(s) (HCPCS)

- 42294913 - Endoscopic hemostatic balloons or needles or tubes or accessories
- 42294928 - Endoscopic probes

Common Procedure Terminology (CPT) Code(s)

- 43227 - Esophagoscopy, flexible, transoral; with control of bleeding, any method
- 43255 - Esophagogastroduodenoscopy, flexible, transoral; with control of bleeding, any method
- 44366 - Small intestinal endoscopy, enteroscopy beyond second portion of duodenum, not including ileum; with control of bleeding (eg, injection, bipolar cautery, unipolar cautery, laser, heater probe, stapler, plasma coagulator)
- 44378 - Small intestinal endoscopy, enteroscopy beyond second portion of duodenum, including ileum; with control of bleeding (eg, injection, bipolar cautery, unipolar cautery, laser, heater probe, stapler, plasma coagulator)
- 44391 - Colonoscopy through stoma; with control of bleeding, any method
- 45334 - Sigmoidoscopy, flexible; with control of bleeding, any method
- 45382 - Colonoscopy, flexible; with control of bleeding, any method

International Classification of Diseases (ICD) Code(s)

- K22.1 - Ulcer of Esophagus (K22.10, K22.11)
- K25 - Gastric Ulcer (K25.0, K25.1, K25.2, K25.3, K25.4, K25.5, K25.6, K25.7, K25.9)
- K26 - Duodenal Ulcer (K26.0, K26.1, K26.2, K26.3, K26.4, K26.5, K26.6, K26.7, K26.9)
- K27 - Peptic Ulcer, site unspecified (K27.0, K27.1, K27.2, K27.3, K27.4, K27.5, K27.6, K27.7, K27.9)
- K28 - Gastrojejunal Ulcer (K28.0, K28.1, K28.2, K28.3, K28.4, K28.5, K28.6, K28.7, K28.9)
- K62.6 - Ulcer off Anus and Rectum
- K63.3 - Ulcer of Intestine

Medicare Severity-Diagnosis Related Group (MS-DRG) Code(s)

- 380 - Complicated peptic ulcer with mcc
- 381 - Complicated peptic ulcer with cc
- 382 - Complicated peptic ulcer without cc/mcc
- 383 - Uncomplicated peptic ulcer with mcc
- 384 - Uncomplicated peptic ulcer without mcc

United Nations Standard Products and Services Code (UNSPSC):

- 42295006 - Endoscopic heater probe units or heater probes
- 42295021 - Endoscopic heater probe unit or heater probe accessories

GLOBAL TRADE ITEM NUMBER (GTIN)

Brighton	
BP-7F-SP-350	00884450088715
BP-10F-SP-350	00884450088692
BP-10F-DP-350	00884450088685
BP-7F-DP-350	00884450088708

CROSS REFERENCE

Merit Medical Endotek				Boston Scientific Gold Probe				COOK Quicksilver			
SKU	French Size	Adapter	Length	SKU	French Size	Adapter	Length	SKU	French Size	Adapter	Length
BP-7F-SP-350	7F	Single Port	350 cm	M00560220	7F	Single Port	350 cm	BCP-7AX	7F	Single Port	350 cm
BP-7F-DP-350	7F	Dual Port	350 cm	Not Offered				BCP-7BX	7F	Dual Port	350 cm
BP-10F-SP-350	10F	Single Port	350 cm	M00560101	10F	Dual Port	300 cm	BCP-10AX	10F	Single Port	350 cm
BP-10F-DP-350	10F	Dual Port	350 cm	Not Offered				BCP-10BX	10F	Dual Port	350 cm

Merit Medical Endotek				Olympus BiCOAG				US Endoscopy Bipolar Hemostasis Probe			
SKU	French Size	Adapter	Length	SKU	French Size	Adapter	Length	SKU	French Size	Adapter	Length
BP-7F-SP-350	7F	Single Port	350 cm	CD-B610LA	7F	Single Port	350 cm	BX00711841	7F	Single Port	350 cm
BP-7F-DP-350	7F	Dual Port	350 cm	CD-B612LA	7F	Dual Port	350 cm	BX00711843	7F	Dual Port	350 cm
BP-10F-SP-350	10F	Single Port	350 cm	CD-B620LA	10F	Single Port	350 cm	BX00711847	10F	Single Port	350 cm
BP-10F-DP-350	10F	Dual Port	350 cm	CD-B622LA	10F	Dual Port	350 cm	Not Offered			

Merit Medical Endotek				ConMed BiCap			
SKU	French Size	Adapter	Length	SKU	French Size	Adapter	Length
BP-7F-SP-350	7F	Single Port	350 cm	BP-7350A	7F	Single Port	350 cm
BP-7F-DP-350	7F	Dual Port	350 cm	Not Offered			
BP-10F-SP-350	10F	Single Port	350 cm	BP-10300A	10F	Single Port	300 cm
BP-10F-DP-350	10F	Dual Port	350 cm	Not Offered			

Merit gathers reimbursement information from third-party sources and presents this information for illustrative purposes only. This information does not constitute reimbursement or legal advice and does not guarantee that this information is accurate, complete, without errors, or that use of any of the codes provided will ensure coverage or payment at any particular level. Medicare may implement policies differently in various parts of the country. Physicians and hospitals should confirm with a particular payor or coding authority, such as the American Medical Association or medical specialty society, which codes or combinations of codes are appropriate for a particular procedure or combination of procedures. Reimbursement for a product or procedure can be different depending upon the setting in which the product is used. Coverage and payment policies also change over time and Merit assumes no obligation to update the information provided herein.

* CPT ©2018 American Medical Association (AMA). All rights reserved. CPT® is a registered trademark of the AMA. Applicable FARS/DFARS restrictions apply to government use. Fee schedules, relative value units, conversion factors and/or related components are not assigned by the AMA, are not part of the CPT, and the AMA is not recommending their use. The AMA does not directly or indirectly practice medicine or dispense medical services. The AMA assumes no liability for data contained or not contained herein.