

SPLASHWire™

HYDROPHILIC GUIDE WIRE

INSTRUCTIONS FOR USE

Carefully read all warnings, precautions and directions prior to use. Failure to do so may result in the improper use of this device which could cause the following complications:

- Shearing of the hydrophilic guide wire
- Release of plastic pieces or fragments from the hydrophilic guide wire which may need to be retrieved from the vasculature.
- Vessel trauma

DESCRIPTION

Merit Medical hydrophilic guide wires are constructed from a high quality, steerable, metallic core wire with a polymer coating. The metallic core wire is utilized throughout the entire length of the wire body. The polymer coating (jacket) extends across the entire length of the guide wire surface. A hydrophilic coating is applied over the radiopaque polymer jacket. The hydrophilic coating extends across the entire length of the guide wire surface. The hydrophilic coating, when activated, provides lubricity across the entire polymer surface allowing the guide wire to navigate through the vasculature. Guide wires are supplied sterile and non-pyrogenic.

INDICATIONS FOR USE

The Merit Hydrophilic Guide Wire is intended to be used in the peripheral vascular system to facilitate the placement of devices during diagnostic and interventional procedures.

CONTRAINDICATIONS

These guide wires are not intended for Percutaneous Transluminal Coronary Angioplasty use.

WARNINGS/ADVERSE REACTIONS

- Inspect wire for damage prior to use, do not use a wire that has been bent, kinked, or damaged. Use of a damaged wire may result in vessel damage or wire fragment release into the vessel.
- Do not reshape the hydrophilic wire by any means. Attempting to reshape the wire may cause damage to the wire.
- Do not manipulate or withdraw the wire through a metal entry needle or a metal dilator, or use this wire with devices which contain metal parts such as atherectomy catheters or laser catheters or metal torque devices. This may result in destruction and /or separation of the outer polyurethane coating requiring retrieval. A plastic entry needle is recommended when using this wire for initial placement, or a catheter, introducer sheath or vessel dilator should replace the needle as soon as the guide wire has been inserted into the vessel.
- Never advance the guide wire against resistance without first determining the reason for the resistance under fluoroscopy. If resistance occurs and the cause of resistance cannot be determined, remove the guide wire and device as a unit. Excessive force against resistance may result in damage to the wire and/ or to the vessel.
- When manipulating, advancing, exchanging, or withdrawing a catheter over the wire, secure and maintain the guide wire in place under fluoroscopy to avoid unexpected guide wire advancement; otherwise damage to the vessel wall by the wire's tip may occur.
- The hydrophilic guide wire should be used only by a physician, who is well trained in manipulation and observation of guide wires under fluoroscopy.

WARNINGS/ADVERSE REACTIONS

- Thrombus
- Emboli
- Arterial or venous vessel wall damage
- Plaque dislodgment
- Hematoma at the puncture site
- Infection
- Vessel perforation
- Vessel spasm
- Hemorrhage
- Vascular thrombosis

- Other potential access site complications leading to bleeding, dissection, or perforation that may require intervention.

PRECAUTIONS

- When using a drug or a device concurrently with the wire, the operator should have a full understanding of the properties/characteristics of the drug or device so as to avoid damage to the hydrophilic guide wire.
- Use care when manipulating this guide wire through a tightened Hemostasis valve.

CAUTIONS

- At least 5 cm of the wire should protrude from the device hub at all times to prevent the wire from sliding entirely into the device due to the low sliding friction of this wire.
- Merit Medical hydrophilic guide wires are packaged in a plastic hoop fitted with a luer hub. This packaging is provided to facilitate compliance with the manufacturer recommended guidelines that the wire must be flushed with saline or heparinized saline prior to use (See instructions for use).
- Contents of unopened, undamaged package are sterile and non-pyrogenic.

PREPARATION FOR USE

1. The surface of the hydrophilic guide wire is not lubricious unless it is wet. Before attempting to remove the guide wire from its' dispenser, inject sterile heparinized saline solution into the luer lock hub end of the dispenser to fill the dispenser coil. This will completely cover the guide wire surface, activate the hydrophilic coating, and make the guide wire very lubricious.

WARNING: Failure to hydrate the dispenser hoop prior to guide wire removal may result in guide wire damage and or difficult removal from the dispenser.

2. After hydrating the guide wire, gently grasp the J-straightener device and pull from the dispenser, once the straightener is separated from the dispenser, continue to remove the wire from the hoop.
3. If guide wire is not properly hydrated, it will be difficult to remove from the dispenser. Inject additional heparinized saline solution into dispenser and repeat step # 2.

PRIOR TO USE

CAUTION

1. **CAUTION:** The safety and effectiveness of the Merit Hydrophilic Guide Wire has not been established in the Coronary or Neuro vasculature.
2. **WARNING:** Preclinical testing with this device shows the potential for clot formation in the absence of anticoagulation. Appropriate anticoagulation therapy should be considered to reduce the potential for thrombus formation on the device.

INSTRUCTIONS FOR USE

1. Fill concurrent device with heparinized saline solution before and during use to ensure smooth movement of the hydrophilic guide wire within the device.
2. Use of sterilized gauze moistened with heparinized saline solution and/or a non metal torque device will facilitate handling of the wire.
3. Insert the guide wire into the device and advance to the desired location.
WARNING: If movement of the wire within the device becomes diminished, remove guide wire and reactivate the hydrophilic coating by wetting its entire surface with a heparinized saline solution.
4. Wipe the guide wire with a 4x4 gauze moistened with heparinized saline solution to remove excess blood from the guide wire surface.
WARNING: Do not use dry gauze as this may damage the guide wire surface resulting in increased resistance when the wire is reinserted into the device.
5. Re-hydrate the guide wire prior to reinsertion into any device or placement into a patient. If additional resistance is felt after re-hydration, replace guide wire.
6. Use of alcohol, antiseptic solutions, or other solvents must be avoided.
WARNING: These solutions may adversely affect the surface of the hydrophilic guide wire.
7. After cleaning the wire, place into the saline filled hoop, proximal end first. The wire may also be placed in a guide wire basin and completely covered with heparinized saline solution.
8. **WARNING:** Hydrophilic guide wires must be kept hydrated through out the entire procedure. Re-hydrate as necessary when the surface starts to dry out.

⚠ ONLY Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.

REUSE PRECAUTION STATEMENT

For single patient use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.



Do not withdraw through metal devices.

	Caution
Rx ONLY	Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.
MD	Medical Device
UDI	Unique Device Identifier
	Do Not Use If Package is Damaged and Consult Instruction for Use
	Use by date: YYYY-MM-DD
	Date of Manufacture: YYYY-MM-DD
REF	Catalog number
LOT	Lot number
	Manufacturer
EC REP	Authorized Representative in the European Community
	Date of Manufacture: YYYY-MM-DD
	Consult Instructions for Use For electronic copy scan QR code, or go to www.merit.com/ifu and enter IFU ID. For printed copy, call U.S.A. or EU Customer Service
	Non-pyrogenic
	Single use
	Do not resterilize
	Keep Dry
	Keep away from sunlight
	Single sterile barrier system
STERILE EO	Sterilized Using Ethylene Oxide



Manufacturer:
 Merit Medical Systems, Inc.
 1600 West Merit Parkway
 South Jordan, Utah 84095 U.S.A.
 1-801-253-1600
 U.S.A. Customer Service 1-800-356-3748



Authorized Representative:
 Merit Medical Ireland Ltd
 Parkmore Business Park West, Galway, Ireland
 EC Customer Service +31 43 3588222