

Vacuum Pressure Syringe

INSTRUCTIONS FOR USE

PRODUCT DESCRIPTION

The VacLok AT™, Vacuum Pressure Syringe, is a general piston syringe constructed using a barrel, plunger, piston seal and locking mechanism. The plunger may be locked in any position along the length of the barrel to maintain a constant vacuum pressure.

INDICATIONS FOR USE

The VacLok AT Vacuum Syringe is used to inject fluids into, or withdraw fluids from the body. It can also be used in cases where a vacuum syringe is preferred (e.g. thrombus, abscess fluid, bile, urine etc.).

CAUTIONS

- Read manufacturer's instructions prior to use
- Rx Only Caution: Federal (USA) law restricts this device to sale by or on the order of a physician trained and/or experienced in the use of the device.
- Inspect the device prior to use to verify that no damage has occurred during shipping.
- Do not use after expiration date.

REUSE PRECAUTION STATEMENT

For single patient use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or reesterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or reesterilization may also create a risk of contamination of the device and/or cause patient infection or crossinfection, including, but not limited to, the transmission of infectious disease(s) from one patient.

CONTRAINDICATIONS

None known

INSTRUCTIONS FOR USE

1. Using aseptic technique, open the package containing the sterile product.
2. Inspect for damage or improper assembly.
3. To activate vacuum pressure (Figure 1):
 - a. Pull plunger back to desired location.
 - a. Twist plunger clockwise until it stops and locks into place.
4. To disconnect vacuum pressure (Figure 2):
 - b. Twist plunger counter-clockwise until plunger stops.
5. To aspirate fluids without locking the syringe, pull plunger back and maintain pressure without turning the plunger.




Figure 1.









Figure 2.

Nominal capacity accuracy = ± 4%.

Rx Only	Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.		Do not use if package is damaged
	Caution: Consult accompanying documents. Read instructions prior to use.		Single use
	Non-pyrogenic		Do not resterilize

Dispose of device per institution protocol

CE 0086



www.merit.com



Manufacturer:
 Merit Medical Systems, Inc.
 1600 West Merit Parkway, South Jordan, Utah 84095 U.S.A.
 1-801-253-1600
 U.S.A. Customer Service 1-800-356-3748



Authorized Representative:
 Merit Medical Ireland Ltd, Parkmore Business Park West, Galway, Ireland
 EC Customer Service +31 43 3588222